



mna
LABORATUVARLARI

Notified Body Number: 2841

AB Tip İnceleme Sertifikası EU Type-Examination Certificate

Belge No / Certificate No : 92-20-03-R02
**Belgelendirme Tarihi - Bir Sonraki Belge Tarihi /
Certification Date / Certificate Validity Date** : 12.03.2021-25.12.2025
Belge Geçerlilik Tarihi / Document Validity Period : 5 yıl / 5 years
**Firma Unvanı ve Adresi /
Company Name and Address** : FAGO MEDİKAL SAN. VE TİC. LTD. ŞTİ.
15 Temmuz Mah. Cami Yolu Cad. No:106 / Z1 Bağcılar/
İSTANBUL

Ürün Adı /Modeller / Product Name / Models : FAGO S 101
Direktifi / Directive : 2016/425 REGULATION
Modülü/Kategori / Module / Category : B MODÜLÜ/ KATEGORİ III
MODULE B / CATEGORY III

Test Rapor No/ları / Test Report No : MNA M-2020-00576, M-2021-00097, M-2021-00383
Ürün Tipi / Product Type:
- EN 149:2001+ A1:2009 Solunumla ilgili koruyucu cihazlar - Parçacıklara karşı koruma amaçlı filtrelili
yarım maskeler/ Respiratory protective devices - Filtering half masks to protect against particles

Ürünün Malzeme Bilgisi / Product Material Information: FAGO S 101 model ürünleri kumaş, kulak kayışı,
burun klipsi ve filtre katmanı kullanılarak imal edilmiştir./ FAGO S 101 model products are manufactured using
fabric, earloop, nose clip and filter layer.
Revizyon nedeni/ Reason for revision: Farklı renkte ürünler eklenmiştir./ Different color products have been
added.

Volkan AKIN
12.03.2021

Karar Verici / Approver

Okan AKEL
12.03.2021

Şirket Müdürü / General manager



MNA Laboratuvarları San. Tic.Ltd .Şti
Adres: Küçükbakkalköy Mahallesi Yenidoğan Cad.No:21 Ataşehir/ İstanbul
Tel: 0216 574 07 08 Faks: 0216 575 13 31 www.mnalab.com







ATTACHMENTS (92-20-03-R02)

To certify the PPE product at Category III level, C2 or D module is accompanied by applying one of the conformity assessment methods along with the EU Type Examination (Module B).

Model : FAGO S 101


PPE SPECIFICATION	PERFORMANCE LEVELS
Classification	FFP2
Reusable / Single Shift Use	NR

PPE produced as a single unit to fit an individual user, all the necessary instructions for manufacturing such PPE on the basis of the approved basic model:

MARKING					
MANUFACTURER: FAGO MEDİKAL SAN. VE TİC. LTD. ŞTİ					
PPE TYPE :					
- EN 149:2001+ A1:2009 Respiratory protective devices - Filtering half masks to protect against particles					
MODEL: FAGO S 101					
PICTOGRAM AND PERFORMANCE LEVELS:					
EN 149:2001+ A1:2009 FFP2 NR					
 NB 2841		 Year Month	 yyyy/mm	 -xx°C +yy°C	 < xx%
Or Condition of Storage					

MNA LABORATORIES SAN. TIC. LTD. ŞTİ declares that the above-mentioned product meets the requirements of the directive according to the EU Directive 2016/425, the safety of the product is covered by the conditions and use specified in this certificate and in the technical file.

ATTACHMENTS (92-20-03-R02)



ATTACHMENTS (92-20-03-R02)



DOCUMENTS IN THE TECHNICAL FILE

- Basic Health Safety Requirements
- Risk Assessment
- Test Reports
- Technical Report

Report No : 92-20-03-R02

Report Date : 12.03.2021

Application No : 92-20-03

1. COMPANY INFORMATION:

FAGO MEDİKAL SAN. VE TİC. LTD. ŐTİ.

15 Temmuz Mah. Cami Yolu Cad. No:106 / Z1 Baęcılar/ İSTANBUL

Tel: +90212 630 67 55 -56

E-mail: info@fagomedikal.com, birsen@fagomedikal.com


2. PPE INFORMATION:

Disposable and non-sterile half mask made of particulate protection filter material.

3. PPE TYPE IDENTIFICATION

EN 149:2001 +A1:2009 Respiratory protective devices - Filtering half masks to protect against particles - Requirements, testing, marking

4. PPE PICTURES





FAGO S 101

5. PPE DIMENSIONS:

FAGO S 101 model has been found to be produced using standard sizes.

6. PPE PRODUCT MATERIAL INFORMATION:

The product is made of elastic strap, nonwoven fabric on the outer and inner layers, filter material on the middle layer.

7. ESSENTIAL HEALTH AND SAFETY REQUIREMENTS

- A visual inspection was made according to EN 149:2001 +A1:2009 for ergonomics.
- Protection levels and degrees are defined by the manufacturer.
- Suitable construction materials were determined by visual inspection according to EN 149:2001 +A1:2009.
- Respiratory protective dimensions are evaluated according to EN 149:2001 +A1:2009.
- Conditioning EN 149:2001 +A1:2009 part 8.3, Penetration EN 149:2001 +A1:2009 part 8.11 (EN 13274-7), Application performance EN 149:2001 +A1:2009 part 8.4, Inward leakage EN 149:2001 +A1:2009 part 8.5, Flammability EN 149:2001 +A1:2009 part 8.6, The carbon dioxide content of the inhaled air EN 149:2001 +A1:2009 part 8.7, Inhalation resistance EN 149:2001 +A1:2009 part 8.9, Exhalation resistance EN 149:2001 +A1:2009 part 8.9 has been tested and evaluated.

8. ANALYSIS AND EVALUATIONS:
EN 149:2001 +A1:2009

TESTS	PARAMETER	PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION
		FFP1	FFP2	FFP3			
Visual inspection	Shall also the marking and the information supplied by the manufacturer				Appropriate	-	PASS
Banned Azo Dyes	< 30 mg/kg				< 5 mg/kg	< 30 mg/kg	PASS
Total inward leakage	At least 46 out of the 50 individual exercise result	<25	<11	<5	See the table below	FFP2	PASS
	At least 8 out of the 10 individual wearer arithmetic means	<22	<8	<2	See the table below	FFP2	PASS

WHITE

Total Inward Leakage (%)						
	Exercise 1	Exercise 2	Exercise 3	Exercise 4	Exercise 5	Average
Subject 1 (As recieved)	4.6	4.9	4.8	5.4	4.7	4.9
Subject 2 (As recieved)	5.4	5.3	4.7	4.8	5.5	5.1
Subject 3 (As recieved)	4.9	5.3	4.9	4.9	4.9	5.0
Subject 4 (As recieved)	4.8	4.9	4.8	5.4	5.5	5.1
Subject 5 (As recieved)	5.4	4.7	4.9	5.0	4.9	5.0
Subject 6 (After temperature conditioning)	4.9	4.9	4.8	5.4	4.8	5.0
Subject 7 (After temperature conditioning)	5.5	5.0	5.1	6.0	6.2	5.6
Subject 8 (After temperature conditioning)	5.5	5.2	5.5	4.7	4.7	5.1
Subject 9 (After temperature conditioning)	4.9	4.8	4.9	4.6	4.9	4.8
Subject 10 (After temperature conditioning)	5.0	4.9	4.7	4.8	4.8	4.8

BLACK

Total Inward Leakage (%)						
	Exercise 1	Exercise 2	Exercise 3	Exercise 4	Exercise 5	Average
Subject 1 (As recieved)	2,6	3,5	1,7	3,0	3,1	2,8
Subject 2 (As recieved)	4,4	3,4	2,6	4,6	2,9	3,6
Subject 3 (As recieved)	4,1	1,7	2,2	2,9	2,8	2,7
Subject 4 (As recieved)	3,8	4,9	2,3	4,6	4,7	4,1
Subject 5 (As recieved)	3,7	4,4	4,2	4,7	4,8	4,4
Subject 6 (After temperature conditioning)	3,5	4,7	4,1	1,8	3,6	3,5
Subject 7 (After temperature conditioning)	3,8	4,1	2,3	2,9	5,1	3,6
Subject 8 (After temperature conditioning)	3,8	4,0	3,7	2,7	3,6	3,6
Subject 9 (After temperature conditioning)	3,9	5,1	3,5	3,6	3,8	4,0
Subject 10 (After temperature conditioning)	2,5	4,8	5,0	4,6	5,2	4,4

TESTS	PARAMETER	PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION								
		FFP1	FFP2	FFP3											
Flammibility	Mask shall not burn or not to continue to burn for more than 5 s				Flame not seen	-	PASS								
Carbondioxide content of the inhalation air	Shall not exceed an average of % 1				<table border="1"> <tr> <th>WHITE</th> <th>BLACK</th> </tr> <tr> <td>0,70</td> <td>0,72</td> </tr> <tr> <td>0,75</td> <td>0,70</td> </tr> <tr> <td>0,71</td> <td>0,71</td> </tr> </table>	WHITE	BLACK	0,70	0,72	0,75	0,70	0,71	0,71	-	PASS
WHITE	BLACK														
0,70	0,72														
0,75	0,70														
0,71	0,71														
Penetration of filter material	Sodium chloride, 95 L/min % max	% 20	% 6	% 1	See the table below	FFP2	PASS								
	Paraffin oil, 95 L/min % max	% 20	% 6	% 1	See the table below	FFP2	PASS								

WHITE

Penetration of filter material	Sodium Chloride (%)	Paraffin Oil (%)
As recieved	3.6	2.8
As recieved	3.3	3.2
As recieved	3.5	3.0
After the simulated wearing treatment	3.2	2.9
After the simulated wearing treatment	3.6	2.6
After the simulated wearing treatment	3.6	3.1
Mechanical strength and temperature conditioning	3.4	3.1
Mechanical strength and temperature conditioning	3.0	3.0
Mechanical strength and temperature conditioning	3.5	3.2

BLACK

Penetration of filter material	Sodium Chloride (%)	Paraffin Oil (%)
As recieved	2,8	2,7
As recieved	2,8	2,7
As recieved	2,8	2,7
After the simulated wearing treatment	2,7	2,8
After the simulated wearing treatment	2,7	2,8
After the simulated wearing treatment	2,9	2,9
Mechanical strength and temperature conditioning	3,0	3,0
Mechanical strength and temperature conditioning	3,0	3,0
Mechanical strength and temperature conditioning	3,1	3,0

TESTS	PARAMETER	PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION
		FFP1	FFP2	FFP3			
Compatibility with skin	Materials shall not be known to be likely to cause irritation or any other adverse effect to health				Appropriate	-	PASS
Head harness	It can be donned and removed easily				Appropriate	-	PASS
Breathing Resistance	Inhalation 30L/min	0,6 mbar	0,7 mbar	1 mbar	See the table below	FFP2	PASS
	Inhalation 95L/min	2,1 mbar	2,4 mbar	3 mbar	See the table below	FFP2	PASS
	Exhalation 160L/min	3 mbar	3 mbar	3 mbar	See the table below	FFP2	PASS

WHITE

Breathing Resistance (mbar)	Inhalation 30L/min (mbar)	Inhalation 95L/min (mbar)
As recieved	0.5	1.9
As recieved	0.5	1.9
As recieved	0.4	1.8
After temperature conditioning	0.4	1.8
After temperature conditioning	0.4	1.9
After temperature conditioning	0.5	1.9
After the simulated wearing treatment	0.4	1.8
After the simulated wearing treatment	0.4	1.9
After the simulated wearing treatment	0.5	1.9

BLACK

Breathing Resistance (mbar)	Inhalation 30L/min (mbar)	Inhalation 95L/min (mbar)
As recieved	0,4	1,7
As recieved	0,4	1,7
As recieved	0,4	1,8
After temperature conditioning	0,4	1,8
After temperature conditioning	0,4	1,8
After temperature conditioning	0,4	1,7
After the simulated wearing treatment	0,5	1,7
After the simulated wearing treatment	0,5	1,7
After the simulated wearing treatment	0,5	1,7

WHITE

Breathing Resistance 160L/min (mbar)	Facing directly ahead	Facing vertically upwards	Facing vertically downwards	Lying on the left side	Lying on the right side
As recieved	2,0	2,0	2,0	2,0	2,0
As recieved	1,9	2,0	2,0	2,0	2,0
As recieved	2,0	2,0	2,0	2,0	2,0
After temperature conditioning	1,9	2,0	1,9	2,0	2,0
After temperature conditioning	1,9	2,0	2,0	1,9	2,0
After temperature conditioning	1,9	2,0	2,0	2,0	2,0
After the simulated wearing treatment	1,9	1,9	1,9	2,0	2,0

After the simulated wearing treatment	2,0	2,0	1,9	1,9	1,9
After the simulated wearing treatment	2,0	2,0	2,0	2,0	2,0

BLACK

Breathing Resistance 160L/min (mbar)	Facing directly ahead	Facing vertically upwards	Facing vertically downwards	Lying on the left side	Lying on the right side
As recieved	1,8	1,8	1,8	1,8	1,9
As recieved	1,8	1,8	1,8	1,8	1,8
As recieved	1,8	1,8	1,8	1,8	1,8
After temperature conditioning	1,8	1,9	1,8	1,8	1,8
After temperature conditioning	1,8	1,9	1,8	1,8	1,8
After temperature conditioning	1,8	1,9	1,8	1,8	1,8
After the simulated wearing treatment	1,8	1,8	1,8	1,8	1,8
After the simulated wearing treatment	1,8	1,8	1,8	1,9	1,9
After the simulated wearing treatment	1,8	1,8	1,8	1,9	1,8

9. DECISION PROPOSAL

Analysis and examinations FAGO S 101 model coded personal protective equipment; Respiratory Protective Devices EN 149:2001 +A1:2009- Filtered Half Masks for Protection Against Particles - Properties, Experiments and Marking standards are evaluated. It is recommended to be certified at the performance levels specified as a result of technical evaluations.

10. ATTACHMENTS

- Basic Health Safety Requirements
- Risk Assessment
- Test Reports
- User Instruction

Reason for revision : Different color products have been added.

CONTROLLER : VOLKAN AKIN

SING :

DATE : 12.03.2021



EU DECLARATION OF CONFORMITY

MANUFACTURER

FAGO MEDİKAL SANAYİ VE TİCARET LİMİTED ŞİRKETİ

15 Temmuz Mahallesi Cami Yolu Caddesi No:106 İç Kapı No: Z1 Bağcılar İSTANBUL /
TURKEY

PRODUCT DESCRIPTION

Brand Name: Fago Model: FAGO S 101

Filtering Half Mask

Class: FFP2 NR

Particle Filtering Half Face Mask in Category III product according to (EU) 2016/425 Personal Protective Equipment Regulation

The Manufacturer declares on his sole responsibility that the product above is, under conditions of normal use and conditions defined by the Manufacturer, safe and meets all the necessary legal conditions and requirements. The product is a personal protective equipment that is intended for single use and solely in accordance with the Manufacturer's instructions.

The Conformity is ensured with the following mechanism:

- Complies with EU 2016/425 Personal Protective Equipment Regulation establishing technical requirements for Category III products,
- Complies with Essential Health and Safety Requirements of Technical harmonised standard EN 149:2001 +A1:2009
- All required tests referred in above standards are conducted,
- Complies with other relevant harmonized legislation and community standards
- For the assessment of conformity the EU Type Examination certificate (Serial No:92-20-03) is issued, after all technical evaluations for conformity to the regulation and harmonised standards conducted, by;
 - MNA LAB SAN TIC LTD STI, as Notified Body number 2841
- The product is under surveillance of same Notified Body, NB 2841 according to the Annex III (Module C2) of the PPE Regulation (EU) 2016/425, for quality assurance.

MARKING, LABELLING

Marking, labelling and user information are prepared in accordance with EU 2016/425 Personal Protective Equipment Regulation and the harmonised product standards given above.

MEASURES TO ENSURE CONFORMITY

The Manufacturer declares that he has taken all necessary measures to ensure the conformity of products placed on the market with technical documentation and technical requirements for this type of product.

GÖKHAN AYDIN

General Manager

25/12/2020

FAGO MEDİKAL SANAYİ VE TİC. LTD. ŞTİ.
15 Temmuz Mh. Cami Yolu Cd. No:106/Z1
Bağcılar/İST. Tic. Sic. No: 240684-5
Günesi V.D.: 384 073 8071
Mersis No.: 0384 0738 0710 0001



2841

EU-KONFORMITÄTSERKLÄRUNG

HERSTELLER

FAGO MEDİKAL SANAYİ VE TİCARET LİMİTED ŞİRKETİ

15 Temmuz Mahallesi Cami Yolu Caddesi Nr.: 106 İç Kapı Nr.: Z1 Bağcılar İSTANBUL / TÜRKİE

PRODUKTBE SCHREIBUNG

Markenname: Fago Modell: FAGO S 101

Halbmaske filtern

Klasse: FFP2 NR

Partikelfilter-Halbgesichtsmaske in Produkt der Kategorie III gemäß (EU) 2016/425 Verordnung über persönliche Schutzausrüstung

Der Hersteller erklärt in seiner alleinigen Verantwortung, dass das oben genannte Produkt unter normalen Verwendungsbedingungen und vom Hersteller festgelegten Bedingungen sicher ist und alle erforderlichen gesetzlichen Bedingungen und Anforderungen erfüllt. Das Produkt ist eine persönliche Schutzausrüstung, die zum einmaligen Gebrauch bestimmt ist und ausschließlich den Anweisungen des Herstellers entspricht.

Die Konformität wird mit folgendem Mechanismus sichergestellt:

- Entspricht der EU-Verordnung über Personenschutz ausrüstung 2016/425 zur Festlegung technischer Anforderungen für Produkte der Kategorie III.
- Entspricht den grundlegenden Gesundheits- und Sicherheitsanforderungen der technischen harmonisierten Norm EN 149: 2001 + A1: 2009
- Alle erforderlichen Tests, auf die in den oben genannten Standards Bezug genommen wird, werden durchgeführt.
- Entspricht anderen relevanten harmonisierten Gesetzen und Gemeinschaftsstandards
- Zur Beurteilung der Konformität wird nach allen technischen Bewertungen der Konformität mit der Verordnung und den harmonisierten Normen das EU-Musterprüfzeugnis (Seriennummer: 92-20-03) ausgestellt. o MNA LAB SAN TIC LTD STI, als benannte Stelle Nummer 2841
- Das Produkt wird zur Qualitätssicherung von derselben benannten Stelle, NB 2841, gemäß Anhang III (Modul C2) der PSA-Verordnung (EU) 2016/425 überwacht.

KENNZEICHNUNG, ETIKETTIERUNG

Kennzeichnung, Kennzeichnung und Benutzerinformationen werden gemäß der EU-Verordnung über persönliche Schutzausrüstung 2016/425 und den oben angegebenen harmonisierten Produktnormen erstellt.

MASSNAHMEN ZUR SICHERHEIT DER KONFORMITÄT

Der Hersteller erklärt, dass er alle erforderlichen Maßnahmen getroffen hat, um die Konformität der in Verkehr gebrachten Produkte mit den technischen Unterlagen und technischen Anforderungen für diesen Produkttyp sicherzustellen.

GÖKHAN AYDIN

General Manager

25/12/2020

FAGO MEDİKAL SAN ve TIC LTD. STI.
15 Temmuz Mh. Cami Yolu Caddesi No: 106/21
Bağcılar/İST. Tic.Sic.No: 244684-5
Güneşli V.D.: 384 073 8071
Mersis No: 0384 0738 0710 0001

CE

2841